



William V. MacGill & Co.
1000 N. Lombard Road
Lombard, IL 60148
(800) 323-2841 / (800) 727-3433 fax
www.macgill.com
macgill@macgill.com

February 18, 2026

URGENT RECALL INFORMATION: MacGill Item #14078 Medline Sterile Saline Wound Wash

Dear Valued MacGill Customer,

Medline has issued a *recall on MDSALINE7 Sterile Saline Wound Wash*

Medline has identified that the manufacturer of the sterile saline wound wash may not have manufactured products to meet the minimum required sterility assurance level required for sterile products. Using the affected product may increase the risk of patient infection, which may result in adverse outcomes and/or require medical intervention, such as prescription antibiotics. There have been no reports received of illness or injury due to this issue. Affected products are to be destroyed.

Please check your stock to see if you received one of the following lot numbers:

92272, 92297, 102254, 102266, 112206, 112225, 12351, 122277, 12378, 22324, 22382, 42313, 52381, 62346, 623100, 72373, 82359, 92366, 102323, 102354, 112399, 122313, 122352, 12414, 12495, 124130, 124170, 224115, 32403, 32459, 42423, 424105, 52438, 524135, 62402, 52478, 624117, 72405, 72466, 724121, 82427, 924103, 924035, 1024001, 1024030, 1124007, 1124033, 0125002, 0125003, 0125070, 0225002, 0225062, 0325002, 0325126, 0425001, 0425085, 0525102, 0625044, 0625119, 0725028, 0825052, 0925076

If you are in possession of an affected lot number please destroy the affected product.

Sincerely,

MacGill's Customer Service Department